



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY  
WASHINGTON, D.C. 20460

FEB 15 2000

Debra Lynch  
Technical Support Chemist  
Zellweger Analytics  
Lachat Instruments Divisions  
6645 West Mill Road  
Milwaukee, WI 53218-1239

OFFICE OF  
WATER

Dear Ms. Lynch:

The Analytical Methods Staff (AMS) has reviewed Lachat QuikChem Method 10-107-06-2-D: "Determination of Total Kjeldahl Nitrogen [TKN] By Flow Injection Analysis Colorimetry" [October 7, 1997] for equivalency to EPA Method 351.2 for the determination of TKN in wastewater (ATP Case No. N97-0002). Based on this review, AMS has determined that these methods are equivalent and that Method 10-107-06-2-D may be used for National Pollutant Discharge Elimination System (NPDES) compliance monitoring provided whole, homogenized samples are analyzed (i.e., samples should not be filtered prior to digestion when results are reported to a permit authority).

We appreciate your interest in the development of environmental monitoring methods. If you have any questions regarding review of this ATP, please contact Maria Gomez-Taylor (202/260-1639) at your convenience.

Sincerely,

A handwritten signature in blue ink, appearing to read "W. A. Telliard".

William A. Telliard, Director  
Analytical Methods Staff  
Engineering and Analysis Division (4303)

cc: USEPA Regional Administrators (all Regions)  
Quality Assurance Managers (all Regions)  
Water Management Division Directors (all Regions)  
ATP Coordinators (all Regions)  
Maria Gomez-Taylor, USEPA, EAD  
Khouane Ditthavong, USEPA, EAD  
James Boiani, DynCorp I&ET