



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
OFFICE OF RESEARCH AND DEVELOPMENT
ENVIRONMENTAL MONITORING SYSTEMS LABORATORY
CINCINNATI, OHIO 45268

February 23, 1995

David Diamond
Application Chemist
Lachat Instruments
6645 West Mill Road
Milwaukee, Wisconsin 53218-1239

Dear Mr. Diamond:

We are pleased to inform you that in the judgment of our technical staff, the proposed method 10-115-01-1-A dated October 14, 1994 is an acceptable version of the approved method EPA 365.1 for drinking water and wastewater, dated 1978 and 1993. Therefore, no alternate test procedure application is necessary. The Alternate Test Procedures Program Staff thanks you for your interest in the development of environmental methods.

Sincerely yours,

James W. Eichelberger
Director
Quality Assurance Research Division

cc: Patrice Hoeschele, VP of Technical Services, Lachat Instruments
Office of Ground Water and Drinking Water (4601), Office of Water, U.S. Environmental
Protection Agency, 401 M Street, S.W., Washington, DC 20460
304H Committee Chairman
USEPA Regional Administrator (All regions)
Quality Assurance Managers (All regions)
Environmental Services Division Chiefs (All regions)
Director, Water Management Division (All regions)

