



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY  
OFFICE OF RESEARCH AND DEVELOPMENT  
ENVIRONMENTAL MONITORING SYSTEMS LABORATORY  
CINCINNATI, OHIO 45268

March 15, 1995

David Diamond  
Application Chemist  
Lachat Instruments  
6645 West Mill Road  
Milwaukee, Wisconsin 53218-1239

Dear Mr. Diamond:

We are pleased to inform you that in the judgment of our technical staff, the proposed Total Phosphorous Method **10-115-01-1-F**, dated October 14, 1994 is an acceptable version of the approved method EPA 365.1 for wastewater, dated 1978. Therefore, no alternate test procedure application is necessary. The Alternate Test Procedures Program Staff thanks you for your interest in the development of environmental methods.

Sincerely yours,

James W. Eichelberger  
Director

Quality Assurance Research Division

cc: Patrice Hoeschele, VP of Technical Services, Lachat Instruments  
Office of Ground Water and Drinking Water (4601), Office of Water, U.S.  
Environmental Protection Agency, 401 M Street, S.W., Washington, DC  
20460  
304H Committee Chairman  
USEPA Regional Administrator (All Regions)  
Quality Assurance Managers (All Regions)  
Environmental Services Division Chiefs (All Regions)  
Director, Water Management Division (All Regions)

