



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY  
WASHINGTON, D.C. 20460

OFFICE OF  
WATER

Lynn Egan  
Applications Chemist  
Lachat Instruments  
P.O. Box 389, Mail Stop #1  
Loveland, CO 80539-0389

JUL 17 2003

Dear Ms. Egan:

The Statistics and Analytical Support Branch (SASB) and the Office of Ground Water and Drinking Water's Technical Support Center (OGWDW/TSC) have completed technical review of the materials submitted in support of the following applications for approval of alternate test procedures (ATPs) for use in compliance monitoring under National Primary Drinking Water Regulations (NPDWR).

Method Number [Revision Date]	Analyte	ATP Case No. (drinking water)
10-302-00-1-A [11/08/2001]	Conductivity	D03-0002
10-302-00-1-B [11/08/2001]	Conductivity	D03-0001

We are pleased to inform you that in the judgement of our technical staff the above methods are acceptable versions of EPA-approved methods for the determination of conductivity in drinking water using a conductivity cell (e.g. Standard Method 2510 B (18<sup>th</sup>, 19<sup>th</sup>, and 20<sup>th</sup> Editions) and ASTM Method D1125-95 A) and may be used for NPDWR compliance monitoring

We appreciate Lachat Instruments' continued interest in the development of environmental compliance monitoring methods. If you have any questions regarding these determinations or review of these ATP applications, please contact William Telliard of SASB (202/566-1061) or Herb Brass of OGWDW/TSC (513/569-7936) at your convenience.

Sincerely,

William A. Telliard  
Director of Analytical Methods  
Engineering and Analysis Division (4303)

Herb Brass, Ph.D.  
Technical Support Center (MS-140)  
Office of Ground Water and Drinking Water

cc:

USEPA Regional Administrators (all Regions)

Quality Assurance Managers (all Regions)

ATP Coordinators (all Regions)

Water Management Division Directors (all Regions)

Gregory J. Carroll, USEPA, OGWDW

Lillian Holmes, USEPA, OGWDW/TSC

Robin K. Oshiro, USEPA, EAD

James Boiani, DynCorp, SCC