



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY  
CINCINNATI, OHIO 45268

Lynn Egan  
legan@hach.com  
Applications Chemist  
Lachat Instruments  
P.O. Box 389, Mail Stop #1  
Loveland, CO 80539-0389

Dear Ms. Egan:

The Office of Ground Water and Drinking Water's Technical Support Center (OGWDW/TSC) has completed technical review of the materials submitted in support of the following applications for approval as alternate test procedures (ATPs) for use in drinking water compliance monitoring.

QuikChem Method Number [Revision Date]	Analyte	ATP Case No.
10-204-00-1-X2 [04/16/2008]	Cyanide	D07-0006
10-115-01-1-V [04/16/2008]	Orthophosphate	D07-0007
10-107-04-1-R [04/16/2008]	Nitrate-Nitrite	D07-0008

We are pleased to inform you that per the terms of the ATP program, the methods listed in the table above are acceptable versions of approved methods for the determination of the analytes listed and may be used for compliance monitoring under National Primary Drinking Water Regulations (NPDWRs). The QuikChem methods listed above and the approved methods employ the same chemistry to determine the regulated analytes.

We appreciate your interest in the development of environmental monitoring methods. If you have any questions regarding the review of these alternate test procedures, please contact Steve Wendelken at: [wendelken.steve@epa.gov](mailto:wendelken.steve@epa.gov) or by telephone at: 513-569-7491.

Sincerely,

ATP Coordinator  
Office of Ground Water and Drinking Water  
Technical Support Center (MS-140)  
26 W. Martin Luther King Dr.  
Cincinnati, Ohio 45219  
Phone: (513) 569-7491  
Fax: (513) 569-7837  
[wendelken.steve@epa.gov](mailto:wendelken.steve@epa.gov)

cc:  
ATP Coordinators (all Regions)  
Laboratory Certification Coordinators (all Regions)  
Gregory J. Carroll, USEPA, OGWDW  
Danielle Carter, CSC, SCC